BLUE DIAMOND® GEL USE + CARE

Cleaning

- For cleaning Blue Diamond® Gel, we recommend using standard operating room cleaning agents. Any cleaning agent residual should be rinsed with clean water. However, we do not recommend long soaks.

Warnings

- Please avoid using sharp instruments that may puncture the outer film. Do not drag pads and positioning devices across sharp edges.
- DO NOT AUTOCLAVE OR ATTEMPT TO STERILZE THIS PRODUCT
- Gel pads should not be exposed to high heat of flame

Warming

- Blue Diamond Gel pads may be warmed in warming cabinets or by immersing in water to make the pads approach normal human body temperature.

Storage

- When storing the pads, we recommend storing them flat whenever possible and storing at room temperature. The gel will match the temperature of storage and should not be exposed to extreme temperatures

Moving Patients

- It is not recommended to move patients while on the pads. Moving patients while on pads may result in tearing or puncturing of pads.

Protocol

- The protocol for using gel positioners in a hospital may vary depending on the specific hospital and the type of procedure being performed. However, a suggested protocol for Blue Diamond Gel pads is available from the factory. It is recommended that a healthcare professional evaluates the patient pre-operatively to determine the use of gel pads or positioners. The patient's condition, the surgical procedure being performed, and the duration of the surgery should be taken into consideration when determining the appropriate use of gel positioners.

Warranty

- All Blue Diamond Polymer products are covered by a two-year manufacturer’s defective replacement policy. If a pad is deemed defective due to the manufacturing process, it will be replaced at no cost.

Questions

- If you have any questions, please contact the distributor from whom the product was purchased or contact us directly.